

Heteropoly Blue Method¹

Method 8186
0.010 to 1.600 mg/L SiO₂ (spectrophotometers)
Powder Pillows
0.01 to 1.60 mg/L SiO₂ (colorimeters)
Scope and application: For boiler and ultrapure water.

¹ Adapted from Standard Methods for the Examination of Water and Wastewater.




Test preparation

Instrument-specific information

Table 1 shows all of the instruments that have the program for this test. The table also shows sample cell and orientation requirements for reagent addition tests, such as powder pillow or bulk reagent tests.

To use the table, select an instrument, then read across to find the applicable information for this test.

Table 1 Instrument-specific information

Instrument	Sample cell orientation	Sample cell
DR 6000 DR 3800 DR 2800 DR 2700 DR 1900	The fill line is to the right.	2495402 
DR 5000 DR 3900	The fill line is toward the user.	
DR 900	The orientation mark is toward the user.	2401906 

Before starting

Install the instrument cap on the DR900 cell holder before ZERO or READ is pushed.

The reaction times in the test procedure are for samples that are at 20 °C (68 °F). If the sample temperature is 10 °C (50 °F), wait 8 minutes for the first (4-minute) reaction time and 2 minutes for the second (1-minute) reaction time. If the sample temperature is 30 °C (86 °F), wait 2 minutes for the first (4-minute) reaction time and 30 seconds for the second (1-minute) reaction time.

To test for very low concentrations, use the ULR rapid liquid method for the best results.

Review the Safety Data Sheets (MSDS/SDS) for the chemicals that are used. Use the recommended personal protective equipment.

Dispose of reacted solutions according to local, state and federal regulations. Refer to the Safety Data Sheets for disposal information for unused reagents. Refer to the environmental, health and safety staff for your facility and/or local regulatory agencies for further disposal information.

Items to collect


Description	Quantity
Amino Acid F Reagent powder pillows, 10-mL	1
Citric acid powder pillows, 10-mL	2
Molybdate 3 Reagent solution	1 mL
Sample cells (For information about sample cells, adapters or light shields, refer to Instrument-specific information on page 1.)	2

Refer to [Consumables and replacement items](#) on page 5 for order information.

Sample collection

- Collect samples in clean plastic bottles with tight-fitting caps. Do not use glass bottles, which will contaminate the sample.
- Analyze the samples as soon as possible for best results.
- If prompt analysis is not possible, keep the samples at or below 6 °C (43 °F) for up to 28 days.
- Let the sample temperature increase to room temperature before analysis.

Powder pillow procedure




1. Start program **651 Silica LR**. For information about sample cells, adapters or light shields, refer to [Instrument-specific information](#) on page 1.

2. **Prepare the blank:** Fill a sample cell with 10 mL of sample.

3. **Prepare the sample:** Fill a second sample cell with 10 mL of sample.

4. Add 14 drops of Molybdate 3 reagent solution to each cell.



5. Swirl to mix.


6. Start the instrument timer. A 4-minute reaction time starts.

7. After the timer expires, add the contents of one Citric Acid Reagent powder pillow to each sample cell.


8. Swirl to mix.




9. Start the instrument timer. A 1-minute reaction time starts.
The destruction of possible phosphate interference occurs during this period.




10. After the timer expires, add the contents of one Amino Acid F Reagent powder pillow to the prepared sample cell.
Blank: The sample without the Amino Acid F Reagent is the blank.




11. Swirl to mix.




12. Start the instrument timer. A 2-minute reaction time starts.
A blue color shows if silica is present.




13. When the timer expires, clean the blank sample cell.




14. Insert the blank into the cell holder.




15. Push **ZERO**. The display shows 0.000 mg/L SiO₂.



16. Clean the prepared sample cell.



17. Insert the prepared sample into the cell holder.



18. Push **READ**. Results show in mg/L SiO₂.

Interferences

Interfering substance	Interference level
Color	Does not interfere when the original sample is used to zero the instrument.
Iron	Large amounts of both ferrous and ferric iron interfere.
Phosphate	Does not interfere at levels less than 50 mg/L PO ₄ ³⁻ . At 60 mg/L PO ₄ ³⁻ , an interference of -2% occurs. At 75 mg/L PO ₄ ³⁻ , the interference is -11%.

Interfering substance	Interference level
Slow reacting forms of silica	Occasionally a sample contains silica which reacts very slowly with molybdate. The nature of these "molybdate-unreactive" forms is not known. A pretreatment with sodium bicarbonate, then sulfuric acid will make these forms reactive to molybdate. The pretreatment is given in <i>Standard Methods for the Examination of Water and Wastewater</i> under Silica-Digestion with Sodium Bicarbonate. A longer reaction time with the sample and the molybdate and acid reagents (before the citric acid is added) can help as an alternative to the bicarbonate pretreatment.
Sulfides	Interfere at all levels.
Turbidity	Small amounts of turbidity do not interfere when the original sample is used to zero the instrument.

Accuracy check

Standard additions method (sample spike)

Use the standard additions method (for applicable instruments) to validate the test procedure, reagents and instrument and to find if there is an interference in the sample.

Items to collect:

- Silica Standard Solution, 25 mg/L SiO₂
 - Pipet, TenSette[®], 0.1–1.0 mL
 - Pipet tips
1. Use the test procedure to measure the concentration of the sample, then keep the (unspiked) sample in the instrument.
 2. Go to the Standard Additions option in the instrument menu.
 3. Select the values for standard concentration, sample volume and spike volumes.
 4. Open the standard solution.
 5. Prepare three spiked samples: use the TenSette pipet to add 0.1 mL, 0.2 mL and 0.3 mL of the standard solution, respectively, to three 10-mL portions of fresh sample. Mix well.
 6. Use the test procedure to measure the concentration of each of the spiked samples. Start with the smallest sample spike. Measure each of the spiked samples in the instrument.
 7. Select **Graph** to compare the expected results to the actual results.

Note: *If the actual results are significantly different from the expected results, make sure that the sample volumes and sample spikes are measured accurately. The sample volumes and sample spikes that are used should agree with the selections in the standard additions menu. If the results are not within acceptable limits, the sample may contain an interference.*

Standard solution method

Use the standard solution method to validate the test procedure, the reagents and the instrument.

Items to collect:

- Silica Standard Solution, 1.00-mg/L SiO₂
1. Use the test procedure to measure the concentration of the standard solution.
 2. Compare the expected result to the actual result.

Note: *The factory calibration can be adjusted slightly with the standard calibration adjust option so that the instrument shows the expected value of the standard solution. The adjusted calibration is then used for all test results. This adjustment can increase the test accuracy when there are small variations in the reagents or instruments.*

Method performance

The method performance data that follows was derived from laboratory tests that were measured on a spectrophotometer during ideal test conditions. Users can get different results under different test conditions.

Program	Standard	Precision (95% confidence interval)	Sensitivity Concentration change per 0.010 Abs change
651	1.000 mg/L SiO ₂	0.990–1.010 mg/L SiO ₂	0.012 mg/L SiO ₂

Summary of method

Silica and phosphate in the sample react with molybdate ion under acidic conditions to form yellow silicomolybdic acid complexes and phosphomolybdic acid complexes. Addition of citric acid destroys the phosphate complexes. An amino acid is then added to reduce the yellow silicomolybdic acid to an intense blue color, which is proportional to the silica concentration. The measurement wavelength is 815 nm for spectrophotometers (DR 1900: 800 nm) or 610 nm for colorimeters.

Consumables and replacement items

Required reagents

Description	Quantity/Test	Unit	Item no.
Silica Reagent Set, low range, includes:	—	100 tests	2459300
Amino Acid F Reagent Powder Pillow, 10 mL	1	100/pkg	2254069
Citric Acid Powder Pillow, 10 mL	2	100/pkg	2106269
Molybdate 3 Reagent Solution	1 mL	50 mL	199526

Recommended standards and apparatus

Description	Unit	Item no.
Silica Standard Solution, 1-mg/L SiO ₂	500 mL	110649
Silica Standard Solution, 25 mg/L as SiO ₂	236 mL	2122531
Water, deionized	4 L	27256

Optional reagents and apparatus

Description	Unit	Item no.
Sodium Bicarbonate	454 g	77601
Sulfuric Acid Solution, 1.00 N	1000 mL	127053
Pipet, TenSette [®] , 0.1–1.0 mL	each	1970001
Pipet tips for TenSette [®] Pipet, 0.1–1.0 mL	50/pkg	2185696
Pipet tips for TenSette [®] Pipet, 0.1–1.0 mL	1000/pkg	2185628



FOR TECHNICAL ASSISTANCE, PRICE INFORMATION AND ORDERING:
In the U.S.A. – Call toll-free 800-227-4224
Outside the U.S.A. – Contact the HACH office or distributor serving you.
On the Worldwide Web – www.hach.com; E-mail – techhelp@hach.com

HACH COMPANY
WORLD HEADQUARTERS
Telephone: (970) 669-3050
FAX: (970) 669-2932