

# Quality Certificate

## Technical Data for Validation of Total Ammonia Chemkey® Reagents

### Quality Certificate

Technical Data for Total Ammonia Chemkeys:

Part Number	9425200
Sensitivity	0.874 Abs / mg/L
Ordinate intersect	0.004 mg/L
Standard Deviation	< 0.04 mg/L at 0.30 mg/L
Accuracy	< ± 8% at 0.30 mg/L
Correlation Coefficient (r <sup>2</sup> )	> 0.990

The detection and the quantitation limits were determined in conformity with USEPA method 40 CFR Appendix B to Part 136, Revision 1.11.

Range (mg/L as NH <sub>3</sub> -N) <sup>1</sup>	0.05 – 1.50
Detection limit	< 0.05 mg/L

<sup>1</sup> With maximum 0.50 mg/L free ammonia

The series of the smallest and largest calibration standards exhibit normal distribution and are outlier- and trend-free.

The calibration gives a linear function.

Results are certified for nine months after manufacture date when stored at 25°C or below.

Result	Precision Confidence Interval (95%)
0.50 mg/L	± 0.03 mg/L
0.90 mg/L	± 0.03 mg/L



**Bennett Vandiver**  
Sr. Manager, Strategic Quality  
Hach



**Darren MacFarland, Ph.D.**  
Director, Chemistry R&D  
Hach



**Be Right™**

HACH  
5600 Lindbergh Drive  
Loveland, CO USA, 80538  
800-227-4224  
[www.hach.com](http://www.hach.com)

HACH  
Willstätterstr 11, 40549  
Düsseldorf, Germany  
(+49) (211) 52880  
[www.hach-lange.com](http://www.hach-lange.com)